

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051024\
 Data File : VN082065.D
 Acq On : 10 May 2024 13:12
 Operator : JC\MD
 Sample : VN0510WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0510WBL01

Quant Time: May 10 23:19:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	123724	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.106	114	265555	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	225690	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	81889	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	109165	52.279	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.560%	
35) Dibromofluoromethane	8.171	113	80542	50.374	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.740%	
50) Toluene-d8	10.571	98	330995	52.711	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.420%	
62) 4-Bromofluorobenzene	12.853	95	103832	41.829	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 83.660%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051024\
Data File : VN082065.D
Acq On : 10 May 2024 13:12
Operator : JC\MD
Sample : VN0510WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0510WBL01

Quant Time: May 10 23:19:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

