

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051024\
 Data File : VN082070.D
 Acq On : 10 May 2024 15:22
 Operator : JC\MD
 Sample : PB160834ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB160834ZHE#02

Quant Time: May 10 23:23:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	126090	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	278564	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	260480	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	98761	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	114760	53.927	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 107.860%		
35) Dibromofluoromethane	8.171	113	84288	50.255	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.520%		
50) Toluene-d8	10.571	98	356352	54.099	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 108.200%		
62) 4-Bromofluorobenzene	12.853	95	119054	45.721	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 91.440%		
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	24502	34.217	ug/l #	79
20) Methylene Chloride	5.277	84	7312	4.009	ug/l #	83
25) 2-Butanone	7.482	43	23462	19.425	ug/l #	82
43) Isopropyl Acetate	8.688	43	278597	50.223	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051024\
Data File : VN082070.D
Acq On : 10 May 2024 15:22
Operator : JC\MD
Sample : PB160834ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB160834ZHE#02

Quant Time: May 10 23:23:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

