

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051024\
 Data File : VN082071.D
 Acq On : 10 May 2024 15:46
 Operator : JC\MD
 Sample : PB160834ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB160834ZHE#03

Quant Time: May 10 23:23:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	128930	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	278090	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	256172	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	92679	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	117158	53.841	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	107.680%	
35) Dibromofluoromethane	8.165	113	85964	51.342	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	102.680%	
50) Toluene-d8	10.571	98	357921	54.430	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	108.860%	
62) 4-Bromofluorobenzene	12.853	95	116824	44.941	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	=	89.880%	
Target Compounds						
						Qvalue
16) Acetone	4.441	43	24210	32.876	ug/l	95
20) Methylene Chloride	5.277	84	9649	5.174	ug/l #	89
25) 2-Butanone	7.482	43	19666	15.923	ug/l	95
43) Isopropyl Acetate	8.688	43	272588	49.224	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051024\
Data File : VN082071.D
Acq On : 10 May 2024 15:46
Operator : JC\MD
Sample : PB160834ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB160834ZHE#03

Quant Time: May 10 23:23:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

