

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051024\
 Data File : VN082072.D
 Acq On : 10 May 2024 16:10
 Operator : JC\MD
 Sample : PB160834ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB160834ZHE#04

Quant Time: May 10 23:24:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	122395	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	265423	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	243202	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	88387	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	111748	54.097	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 108.200%		
35) Dibromofluoromethane	8.165	113	81117	50.759	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.520%		
50) Toluene-d8	10.565	98	341275	54.375	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 108.740%		
62) 4-Bromofluorobenzene	12.847	95	111613	44.986	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 89.980%		
Target Compounds						
						Qvalue
16) Acetone	4.436	43	31217	46.657	ug/l	95
20) Methylene Chloride	5.277	84	10829	6.117	ug/l	85
37) Ethyl Acetate	7.559	43	4954	1.762	ug/l #	87
43) Isopropyl Acetate	8.688	43	301288	57.003	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051024\
Data File : VN082072.D
Acq On : 10 May 2024 16:10
Operator : JC\MD
Sample : PB160834ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB160834ZHE#04

Quant Time: May 10 23:24:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

