

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051024\
 Data File : VN082077.D
 Acq On : 10 May 2024 18:10
 Operator : JC\MD
 Sample : P2422-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WCS-TP-F

Quant Time: May 10 23:26:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	117820	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	261351	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	240472	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	87927	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	110425	55.532	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.060%
35) Dibromofluoromethane	8.171	113	76217	48.436	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.880%
50) Toluene-d8	10.565	98	330249	53.438	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.880%
62) 4-Bromofluorobenzene	12.853	95	111653	45.703	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	91.400%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	37171	59.038	ug/l	97
20) Methylene Chloride	5.283	84	10117	5.937	ug/l #	94
37) Ethyl Acetate	7.559	43	4564	1.649	ug/l #	88
43) Isopropyl Acetate	8.688	43	361925	69.542	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051024\
Data File : VN082077.D
Acq On : 10 May 2024 18:10
Operator : JC\MD
Sample : P2422-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WCS-TP-F

Quant Time: May 10 23:26:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

