

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051024\
 Data File : VN082078.D
 Acq On : 10 May 2024 18:33
 Operator : JC\MD
 Sample : P2423-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-GG

Quant Time: May 10 23:27:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	120459	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	267792	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.864	117	248363	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	90555	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	110919	54.558	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.120%
35) Dibromofluoromethane	8.165	113	78479	48.674	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.340%
50) Toluene-d8	10.570	98	343513	54.247	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.500%
62) 4-Bromofluorobenzene	12.847	95	113929	45.513	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	91.020%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	38332	59.596	ug/l	98
20) Methylene Chloride	5.283	84	9542	5.477	ug/l #	86
43) Isopropyl Acetate	8.688	43	298391	55.955	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051024\
Data File : VN082078.D
Acq On : 10 May 2024 18:33
Operator : JC\MD
Sample : P2423-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-GG

Quant Time: May 10 23:27:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

