

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051024\
 Data File : VN082082.D
 Acq On : 10 May 2024 20:09
 Operator : JC\MD
 Sample : P2430-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1011UMR-SS003-1224-01

Quant Time: May 10 23:28:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	128715	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.100	114	287074	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	254040	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	94442	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	117695	54.178	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	=	108.360%	
35) Dibromofluoromethane	8.171	113	84194	48.711	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	=	97.420%	
50) Toluene-d8	10.571	98	364418	53.683	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	=	107.360%	
62) 4-Bromofluorobenzene	12.847	95	117706	43.864	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery	=	87.720%	
Target Compounds						
16) Acetone	4.424	43	27275	37.818	ug/l	# 85
20) Methylene Chloride	5.277	84	10321	5.544	ug/l	# 96
43) Isopropyl Acetate	8.688	43	176076	30.801	ug/l	# 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051024\
Data File : VN082082.D
Acq On : 10 May 2024 20:09
Operator : JC\MD
Sample : P2430-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1011UMR-SS003-1224-01

Quant Time: May 10 23:28:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

