

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051024\
 Data File : VN082083.D
 Acq On : 10 May 2024 20:33
 Operator : JC\MD
 Sample : PB160862TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB160862TB

Quant Time: May 10 23:29:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	121418	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	259169	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	240188	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	86270	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	110922	54.129	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.260%
35) Dibromofluoromethane	8.165	113	77764	49.835	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.680%
50) Toluene-d8	10.565	98	340477	55.557	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	111.120%
62) 4-Bromofluorobenzene	12.853	95	111532	46.038	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	92.080%
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	31337	47.280	ug/l	97
20) Methylene Chloride	5.277	84	10555	6.010	ug/l #	77
43) Isopropyl Acetate	8.689	43	246305	47.725	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051024\
Data File : VN082083.D
Acq On : 10 May 2024 20:33
Operator : JC\MD
Sample : PB160862TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB160862TB

Quant Time: May 10 23:29:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

