

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051120\
 Data File : VN061384.D
 Acq On : 11 May 2020 12:59
 Operator : JC/MD
 Sample : L2525-01ME
 Misc : 2.58g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DRUM-1ME

Quant Time: May 12 09:00:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 05 07:01:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	95974	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	166654	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	162097	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	80408	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	75434	49.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.34%	
35) Dibromofluoromethane	7.54	113	41102	41.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	83.42%	
50) Toluene-d8	10.06	98	210479	52.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.78%	
62) 4-Bromofluorobenzene	12.38	95	80432	53.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.40%	
Target Compounds						
18) Methyl Acetate	4.27	43	12474	7.098	ug/l	95
31) Cyclohexane	7.61	56	14368	6.504	ug/l #	90
39) Methylcyclohexane	9.04	83	4038	2.158	ug/l	94
69) o-Xylene	11.92	106	4512	2.132	ug/l	90
80) 1,3,5-Trimethylbenzene	12.71	105	32160	6.573	ug/l	100
84) 1,2,4-Trimethylbenzene	13.02	105	5438	1.079	ug/l	97
86) p-Isopropyltoluene	13.26	119	7329	1.446	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051120\
Data File : VN061384.D
Acq On : 11 May 2020 12:59
Operator : JC/MD
Sample : L2525-01ME
Misc : 2.58g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DRUM-1ME

Quant Time: May 12 09:00:14 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Tue May 05 07:01:08 2020
Response via : Initial Calibration

