

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051120\
 Data File : VN061388.D
 Acq On : 11 May 2020 15:45
 Operator : JC/MD
 Sample : L2544-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SS037MW218-200505

Manual Integrations
 APPROVED

sam
 5/12/2020 3:34:30 PM

Quant Time: May 12 09:16:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 05 07:01:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	110248	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	196796	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	187487	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	80801	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	89561	51.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.68%	
35) Dibromofluoromethane	7.55	113	59451	51.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.18%	
50) Toluene-d8	10.06	98	245662	51.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.58%	
62) 4-Bromofluorobenzene	12.38	95	91475	51.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.48%	
Target Compounds						
4) Vinyl Chloride	2.16	62	5303m	3.297	ug/l	
27) cis-1,2-Dichloroethene	6.78	96	16597	11.166	ug/l	89
65) Chlorobenzene	11.41	112	7852	2.111	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051120\
Data File : VN061388.D
Acq On : 11 May 2020 15:45
Operator : JC/MD
Sample : L2544-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SS037MW218-200505

Manual Integrations
APPROVED

sam
5/12/2020 3:34:30 PM

Quant Time: May 12 09:16:32 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Tue May 05 07:01:08 2020
Response via : Initial Calibration

