

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051122\
 Data File : VN072410.D
 Acq On : 11 May 2022 12:27
 Operator : JC\MD
 Sample : N2770-03RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-15842RE

Quant Time: May 12 02:10:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 10 15:43:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	211913	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	381264	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	435995	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	181636	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	160763	49.529	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.060%
35) Dibromofluoromethane	8.022	113	122513	48.299	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.600%
50) Toluene-d8	10.439	98	494120	51.149	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.300%
62) 4-Bromofluorobenzene	12.727	95	205606	64.086	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	128.180%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.357	59	21576	40.653	ug/l #	88
16) Acetone	4.287	43	37775	29.387	ug/l	94
25) 2-Butanone	7.334	43	5380	2.652	ug/l #	86
59) 2-Hexanone	11.086	43	8727	2.985	ug/l	62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051122\
Data File : VN072410.D
Acq On : 11 May 2022 12:27
Operator : JC\MD
Sample : N2770-03RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-15842RE

Quant Time: May 12 02:10:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
Quant Title : SW846 8260
QLast Update : Tue May 10 15:43:28 2022
Response via : Initial Calibration

