

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051122\
 Data File : VN072412.D
 Acq On : 11 May 2022 13:15
 Operator : JC\MD
 Sample : N2763-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-4-20220509-A

Quant Time: May 12 02:11:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 10 15:43:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	215201	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	403891	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	445302	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	184174	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	167573	50.839	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	101.680%	
35) Dibromofluoromethane	8.016	113	130503	48.567	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	97.140%	
50) Toluene-d8	10.439	98	510767	49.910	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	99.820%	
62) 4-Bromofluorobenzene	12.727	95	206657	60.805	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	121.620%	
Target Compounds						
16) Acetone	4.310	43	2461	1.885	ug/l	Qvalue # 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051122\
Data File : VN072412.D
Acq On : 11 May 2022 13:15
Operator : JC\MD
Sample : N2763-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-4-20220509-A

Quant Time: May 12 02:11:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
Quant Title : SW846 8260
QLast Update : Tue May 10 15:43:28 2022
Response via : Initial Calibration

