

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN051122\  
 Data File : VN072415.D  
 Acq On : 11 May 2022 14:27  
 Operator : JC\MD  
 Sample : N2763-07  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 MW-1-20220509

Quant Time: May 12 02:12:04 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N051022W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue May 10 15:43:28 2022  
 Response via : Initial Calibration

| Compound                    | R.T.           | QIon | Response   | Conc     | Units  | Dev(Min) |
|-----------------------------|----------------|------|------------|----------|--------|----------|
| -----                       |                |      |            |          |        |          |
| Internal Standards          |                |      |            |          |        |          |
| 1) Pentafluorobenzene       | 8.081          | 168  | 218479     | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 8.963          | 114  | 401165     | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 11.739         | 117  | 438869     | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.669         | 152  | 192626     | 50.000   | ug/l   | 0.00     |
|                             |                |      |            |          |        |          |
| System Monitoring Compounds |                |      |            |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 8.434          | 65   | 166577     | 49.778   | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 61 - 141 |      | Recovery = | 99.560%  |        |          |
| 35) Dibromofluoromethane    | 8.022          | 113  | 127758     | 47.869   | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 69 - 133 |      | Recovery = | 95.740%  |        |          |
| 50) Toluene-d8              | 10.439         | 98   | 498430     | 49.035   | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 65 - 126 |      | Recovery = | 98.080%  |        |          |
| 62) 4-Bromofluorobenzene    | 12.727         | 95   | 206420     | 61.148   | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 58 - 135 |      | Recovery = | 122.300% |        |          |
|                             |                |      |            |          |        |          |
| Target Compounds            |                |      |            |          | Qvalue |          |
| 16) Acetone                 | 4.287          | 43   | 50954      | 38.448   | ug/l # | 83       |
| 17) Carbon Disulfide        | 4.546          | 76   | 5129       | 0.900    | ug/l # | 88       |
| 25) 2-Butanone              | 7.345          | 43   | 13936      | 6.663    | ug/l # | 86       |
| 40) Benzene                 | 8.457          | 78   | 27857      | 2.433    | ug/l   | 95       |
| 52) Toluene                 | 10.504         | 92   | 3115       | 0.441    | ug/l   | 83       |
| 86) p-Isopropyltoluene      | 13.610         | 119  | 4331       | 0.364    | ug/l   | 93       |
| -----                       |                |      |            |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN051122\  
Data File : VN072415.D  
Acq On : 11 May 2022 14:27  
Operator : JC\MD  
Sample : N2763-07  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
MW-1-20220509

Quant Time: May 12 02:12:04 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N051022W.M  
Quant Title : SW846 8260  
QLast Update : Tue May 10 15:43:28 2022  
Response via : Initial Calibration

