

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051122\
 Data File : VN072416.D
 Acq On : 11 May 2022 14:51
 Operator : JC\MD
 Sample : N2763-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-20220509

Quant Time: May 12 02:12:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 10 15:43:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	216959	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	397823	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	440742	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	181929	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	167372	50.366	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	100.740%
35) Dibromofluoromethane	8.016	113	127657	48.232	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.460%
50) Toluene-d8	10.439	98	498564	49.461	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.920%
62) 4-Bromofluorobenzene	12.727	95	207977	62.127	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	124.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051122\
Data File : VN072416.D
Acq On : 11 May 2022 14:51
Operator : JC\MD
Sample : N2763-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-20220509

Quant Time: May 12 02:12:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
Quant Title : SW846 8260
QLast Update : Tue May 10 15:43:28 2022
Response via : Initial Calibration

