

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051122\
 Data File : VN072418.D
 Acq On : 11 May 2022 15:40
 Operator : JC\MD
 Sample : N2782-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RB-220510

Quant Time: May 12 02:12:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 10 15:43:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	205510	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	379497	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	415704	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	162035	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	161014	51.152	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.300%
35) Dibromofluoromethane	8.022	113	120488	47.722	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.440%
50) Toluene-d8	10.439	98	476547	49.559	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.120%
62) 4-Bromofluorobenzene	12.727	95	188658	59.077	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	118.160%
Target Compounds						
16) Acetone	4.298	43	2967	2.380	ug/l	99
29) Tetrahydrofuran	7.686	42	5298	4.143	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051122\
Data File : VN072418.D
Acq On : 11 May 2022 15:40
Operator : JC\MD
Sample : N2782-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RB-220510

Quant Time: May 12 02:12:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
Quant Title : SW846 8260
QLast Update : Tue May 10 15:43:28 2022
Response via : Initial Calibration

