

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051122\
 Data File : VN072419.D
 Acq On : 11 May 2022 16:04
 Operator : JC\MD
 Sample : N2791-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-2022-05-11

Quant Time: May 12 02:12:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 10 15:43:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	205021	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	375178	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	422009	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	169787	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	158621	50.512	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.020%
35) Dibromofluoromethane	8.022	113	121262	48.582	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.160%
50) Toluene-d8	10.439	98	472143	49.667	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.340%
62) 4-Bromofluorobenzene	12.727	95	193312	61.232	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	122.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051122\
Data File : VN072419.D
Acq On : 11 May 2022 16:04
Operator : JC\MD
Sample : N2791-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-2022-05-11

Quant Time: May 12 02:12:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
Quant Title : SW846 8260
QLast Update : Tue May 10 15:43:28 2022
Response via : Initial Calibration

