

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051122\
 Data File : VN072421.D
 Acq On : 11 May 2022 16:52
 Operator : JC\MD
 Sample : N2763-09RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-20220509RE

Quant Time: May 12 02:13:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 10 15:43:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	204757	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	379964	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	425413	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	173106	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	163802	52.229	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.460%	
35) Dibromofluoromethane	8.022	113	123166	48.723	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 97.440%	
50) Toluene-d8	10.439	98	487163	50.601	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 101.200%	
62) 4-Bromofluorobenzene	12.727	95	199084	62.266	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 124.540%	
Target Compounds						Qvalue
16) Acetone	4.293	43	2828	2.277	ug/l	# 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051122\
Data File : VN072421.D
Acq On : 11 May 2022 16:52
Operator : JC\MD
Sample : N2763-09RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-20220509RE

Quant Time: May 12 02:13:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
Quant Title : SW846 8260
QLast Update : Tue May 10 15:43:28 2022
Response via : Initial Calibration

