

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051122\
 Data File : VN072437.D
 Acq On : 12 May 2022 00:31
 Operator : JC\MD
 Sample : N2800-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-11

Quant Time: May 12 02:19:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 10 15:43:28 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	193982	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	353790	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	380626	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	153810	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	150560	50.674	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.340%
35) Dibromofluoromethane	8.022	113	113247	48.113	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.220%
50) Toluene-d8	10.439	98	443993	49.529	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.060%
62) 4-Bromofluorobenzene	12.727	95	172846	58.059	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	116.120%
Target Compounds						
						Qvalue
31) Cyclohexane	8.086	56	5779	1.468	ug/l	# 45
39) Methylcyclohexane	9.457	83	2276	0.599	ug/l	95
59) 2-Hexanone	11.080	43	3301	1.217	ug/l	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051122\
Data File : VN072437.D
Acq On : 12 May 2022 00:31
Operator : JC\MD
Sample : N2800-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-11

Quant Time: May 12 02:19:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
Quant Title : SW846 8260
QLast Update : Tue May 10 15:43:28 2022
Response via : Initial Calibration

