


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051122\
Data File : VN072405.D
Acq On : 11 May 2022 09:46
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
Title : SW846 8260
Last Update : Tue May 10 15:43:28 2022



AutoFind: Scans 1825, 1826, 1827; Background Corrected with Scan 1818

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.1	8128	PASS
75	95	30	60	53.5	20616	PASS
95	95	100	100	100.0	38520	PASS
96	95	5	9	6.2	2369	PASS
173	174	0.00	2	1.4	437	PASS
174	95	50	100	79.4	30603	PASS
175	174	5	9	7.3	2236	PASS
176	174	95	101	96.1	29411	PASS
177	176	5	9	6.4	1872	PASS