

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051222\
 Data File : VN072449.D
 Acq On : 12 May 2022 13:43
 Operator : JC\MD
 Sample : N2798-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DEWATERING-EFF-051022

Quant Time: May 13 02:43:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 10 15:43:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	197949	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	360374	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	415302	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	187813	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	154018	50.799	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.600%
35) Dibromofluoromethane	8.016	113	116067	48.411	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.820%
50) Toluene-d8	10.439	98	447852	49.047	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.100%
62) 4-Bromofluorobenzene	12.727	95	201626	66.489	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	132.980%
Target Compounds						
29) Tetrahydrofuran	7.692	42	7862	6.383	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051222\
Data File : VN072449.D
Acq On : 12 May 2022 13:43
Operator : JC\MD
Sample : N2798-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DEWATERING-EFF-051022

Quant Time: May 13 02:43:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
Quant Title : SW846 8260
QLast Update : Tue May 10 15:43:28 2022
Response via : Initial Calibration

