

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051222\
 Data File : VN072452.D
 Acq On : 12 May 2022 14:55
 Operator : JC\MD
 Sample : N2800-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-11

Quant Time: May 13 02:44:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 10 15:43:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	191795	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.969	114	304611	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	261583	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	95415	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	176477	60.074	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 120.140%	
35) Dibromofluoromethane	8.022	113	120470	59.445	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 118.900%	
50) Toluene-d8	10.439	98	417033	54.032	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 108.060%	
62) 4-Bromofluorobenzene	12.727	95	127003	49.548	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 99.100%	
Target Compounds						
					Qvalue	
39) Methylcyclohexane	9.451	83	2914	0.891	ug/l	# 61
67) Ethyl Benzene	11.845	91	7479	0.811	ug/l	# 88
78) n-propylbenzene	12.915	91	4076	0.482	ug/l	99
80) 1,3,5-Trimethylbenzene	13.062	105	6733	1.040	ug/l	98
84) 1,2,4-Trimethylbenzene	13.368	105	3994	0.634	ug/l	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051222\
Data File : VN072452.D
Acq On : 12 May 2022 14:55
Operator : JC\MD
Sample : N2800-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-11

Quant Time: May 13 02:44:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
Quant Title : SW846 8260
QLast Update : Tue May 10 15:43:28 2022
Response via : Initial Calibration

