

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051222\
 Data File : VN072454.D
 Acq On : 12 May 2022 15:44
 Operator : JC\MD
 Sample : N2800-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-14

Quant Time: May 13 02:45:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 10 15:43:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	172678	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	283412	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	242703	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	84286	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	164600	62.234	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 124.460%	
35) Dibromofluoromethane	8.016	113	109902	58.287	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 116.580%	
50) Toluene-d8	10.439	98	376519	52.432	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 104.860%	
62) 4-Bromofluorobenzene	12.727	95	110819	46.468	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 92.940%	
Target Compounds						
30) Chloroform	7.816	83	39191	10.384	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051222\
Data File : VN072454.D
Acq On : 12 May 2022 15:44
Operator : JC\MD
Sample : N2800-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-14

Quant Time: May 13 02:45:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
Quant Title : SW846 8260
QLast Update : Tue May 10 15:43:28 2022
Response via : Initial Calibration

