

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051222\
 Data File : VN072464.D
 Acq On : 12 May 2022 19:45
 Operator : JC\MD
 Sample : PB144752TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB144752TB

Quant Time: May 13 02:49:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 10 15:43:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	217950	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	364763	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	327259	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	118118	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	178195	53.379	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 106.760%	
35) Dibromofluoromethane	8.022	113	122557	50.502	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.000%	
50) Toluene-d8	10.439	98	460245	49.797	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 99.600%	
62) 4-Bromofluorobenzene	12.727	95	146923	47.867	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 95.740%	
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	38982	29.486	ug/l #	86
20) Methylene Chloride	5.087	84	26815	9.892	ug/l	88
43) Isopropyl Acetate	8.557	43	27728	4.317	ug/l #	90
84) 1,2,4-Trimethylbenzene	13.363	105	15543	1.994	ug/l	91
95) Naphthalene	15.498	128	26169	7.518	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051222\
Data File : VN072464.D
Acq On : 12 May 2022 19:45
Operator : JC\MD
Sample : PB144752TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB144752TB

Quant Time: May 13 02:49:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
Quant Title : SW846 8260
QLast Update : Tue May 10 15:43:28 2022
Response via : Initial Calibration

