

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051222\
 Data File : VN072465.D
 Acq On : 12 May 2022 20:09
 Operator : JC\MD
 Sample : PB144752ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB144752ZHE#01

Quant Time: May 13 02:49:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 10 15:43:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	219995	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.969	114	407037	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	465721	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	202993	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	166674	49.464	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.920%
35) Dibromofluoromethane	8.022	113	124361	45.924	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	91.840%
50) Toluene-d8	10.439	98	523990	50.806	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.620%
62) 4-Bromofluorobenzene	12.727	95	222425	64.939	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	129.880%
Target Compounds						
					Qvalue	
16) Acetone	4.298	43	38338	28.729	ug/l	93
20) Methylene Chloride	5.093	84	25449	9.301	ug/l	90
43) Isopropyl Acetate	8.569	43	47161	6.580	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051222\
Data File : VN072465.D
Acq On : 12 May 2022 20:09
Operator : JC\MD
Sample : PB144752ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB144752ZHE#01

Quant Time: May 13 02:49:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
Quant Title : SW846 8260
QLast Update : Tue May 10 15:43:28 2022
Response via : Initial Calibration

