

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051222\
 Data File : VN072467.D
 Acq On : 12 May 2022 20:58
 Operator : JC\MD
 Sample : N2760-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-34

Quant Time: May 13 02:50:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 10 15:43:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	242355	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	400300	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	323830	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	124655	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	183290	49.377	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.760%
35) Dibromofluoromethane	8.022	113	127545	47.892	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.780%
50) Toluene-d8	10.439	98	451914	44.555	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	89.120%
62) 4-Bromofluorobenzene	12.727	95	134755	40.005	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	80.000%
Target Compounds						
					Qvalue	
16) Acetone	4.292	43	40099	27.276	ug/l #	82
18) Methyl Acetate	4.869	43	5596	1.293	ug/l	91
20) Methylene Chloride	5.104	84	24517	8.134	ug/l	95
43) Isopropyl Acetate	8.569	43	65092	9.234	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051222\
Data File : VN072467.D
Acq On : 12 May 2022 20:58
Operator : JC\MD
Sample : N2760-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-34

Quant Time: May 13 02:50:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
Quant Title : SW846 8260
QLast Update : Tue May 10 15:43:28 2022
Response via : Initial Calibration

