

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051223\
 Data File : VN077669.D
 Acq On : 12 May 2023 13:12
 Operator : JC\MD
 Sample : 02752-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 230510034-01-TRIP-BLANK

Quant Time: May 15 01:37:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N050823W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue May 09 05:58:09 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/15/2023
 Supervised By : Mahesh Dadoda 05/15/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.818	128	124783	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.106	114	711396	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.870	117	625527	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.582	65	255971	28.525	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	95.067%
60) 4-Bromofluorobenzene	12.853	95	252483m	25.385	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	84.600%
63) Toluene-d8	10.571	98	847559	29.336	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	97.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051223\
Data File : VN077669.D
Acq On : 12 May 2023 13:12
Operator : JC\MD
Sample : 02752-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
230510034-01-TRIP-BLANK

Quant Time: May 15 01:37:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N050823W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue May 09 05:58:09 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/15/2023
Supervised By :Mahesh Dadoda 05/15/2023

