

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051223\
 Data File : VN077671.D
 Acq On : 12 May 2023 14:32
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: May 15 01:37:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N050823W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue May 09 05:58:09 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.818	128	102784	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.106	114	595827	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	521257	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.582	65	213886	28.936	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	96.467%
60) 4-Bromofluorobenzene	12.859	95	196632	23.724	ug/l	0.01
Spiked Amount	30.000	Range	63 - 112	Recovery	=	79.067%
63) Toluene-d8	10.571	98	693624	28.811	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	96.033%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051223\
Data File : VN077671.D
Acq On : 12 May 2023 14:32
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: May 15 01:37:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N050823W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue May 09 05:58:09 2023
Response via : Initial Calibration

