

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051225\
 Data File : VN086571.D
 Acq On : 12 May 2025 11:17
 Operator : JC\MD
 Sample : VN0512MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0512MBL01

Quant Time: May 13 01:47:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	168085	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	307188	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	291005	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	119497	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	113933	46.744	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.480%
35) Dibromofluoromethane	8.165	113	93148	65.335	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	130.660%#
50) Toluene-d8	10.565	98	393629	51.660	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.320%
62) 4-Bromofluorobenzene	12.847	95	133602	48.072	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051225\
Data File : VN086571.D
Acq On : 12 May 2025 11:17
Operator : JC\MD
Sample : VN0512MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0512MBL01

Quant Time: May 13 01:47:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

