

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051225\
 Data File : VN086575.D
 Acq On : 12 May 2025 12:51
 Operator : JC\MD
 Sample : PB167932TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167932TB

Quant Time: May 13 01:50:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	164154	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	304586	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	290524	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	117340	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	112541	47.279	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.560%
35) Dibromofluoromethane	8.165	113	91051	64.410	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	128.820%#
50) Toluene-d8	10.565	98	391528	51.823	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.640%
62) 4-Bromofluorobenzene	12.847	95	135105	49.028	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	98.060%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	22102	21.643	ug/l	99
20) Methylene Chloride	5.271	84	5121	2.327	ug/l	89
43) Isopropyl Acetate	8.688	43	24128	2.710	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051225\
Data File : VN086575.D
Acq On : 12 May 2025 12:51
Operator : JC\MD
Sample : PB167932TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167932TB

Quant Time: May 13 01:50:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

