

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051225\
 Data File : VN086583.D
 Acq On : 12 May 2025 16:05
 Operator : JC\MD
 Sample : PB167932ZHE#08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167932ZHE#08

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 05/13/2025
 Supervised By :Semsettin Yesilyurt 05/13/2025

Quant Time: May 13 01:54:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	180857	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	340176	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	325658	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	136621	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	126519	48.242	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	96.480%		
35) Dibromofluoromethane	8.165	113	101367	64.205	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	128.420%#		
50) Toluene-d8	10.565	98	432991	51.315	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	102.640%		
62) 4-Bromofluorobenzene	12.847	95	153326	49.819	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	99.640%		
Target Compounds					Qvalue	
16) Acetone	4.430	43	22171	19.182	ug/l	99
20) Methylene Chloride	5.283	84	4405	1.817	ug/l #	82
43) Isopropyl Acetate	8.688	43	25078m	2.392	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051225\
Data File : VN086583.D
Acq On : 12 May 2025 16:05
Operator : JC\MD
Sample : PB167932ZHE#08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167932ZHE#08

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 05/13/2025
Supervised By :Semsettin Yesilyurt 05/13/2025

Quant Time: May 13 01:54:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

