

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051319\
 Data File : VN055580.D
 Acq On : 13 May 2019 17:26
 Operator : JC/SP
 Sample : K2840-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 KR-02 4-14

Quant Time: May 14 03:22:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N050819W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 10 06:48:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	959794	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1579199	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1324987	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	391586	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	609095	42.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.30%	
35) Dibromofluoromethane	7.59	113	483186	47.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.86%	
50) Toluene-d8	10.09	98	1906446	48.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.04%	
62) 4-Bromofluorobenzene	12.40	95	562720	42.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.08%	
Target Compounds						
16) Acetone	3.82	43	8847	1.322	ug/l #	86
64) Tetrachloroethene	10.63	164	11832	1.090	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN051319\
Data File : VN055580.D
Acq On : 13 May 2019 17:26
Operator : JC/SP
Sample : K2840-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
KR-02_4-14

Quant Time: May 14 03:22:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N050819W.M
Quant Title : SW846 8260
QLast Update : Fri May 10 06:48:32 2019
Response via : Initial Calibration

