

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN051319\
 Data File : VN055567.D
 Acq On : 13 May 2019 10:04
 Operator : JC/SP
 Sample : VN0513MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0513MBL01

Quant Time: May 14 02:01:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N050819W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 10 06:48:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	789356	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1321316	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1082583	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	272641	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	570202	47.98	ug/l	0.00
Spiked Amount						
			Recovery	=	95.96%	
35) Dibromofluoromethane	7.59	113	451263	52.94	ug/l	0.00
Spiked Amount						
			Recovery	=	105.88%	
50) Toluene-d8	10.09	98	1709747	51.47	ug/l	0.00
Spiked Amount						
			Recovery	=	102.94%	
62) 4-Bromofluorobenzene	12.40	95	468688	41.85	ug/l	0.00
Spiked Amount						
			Recovery	=	83.70%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN051319\
Data File : VN055567.D
Acq On : 13 May 2019 10:04
Operator : JC/SP
Sample : VN0513MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0513MBL01

Quant Time: May 14 02:01:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N050819W.M
Quant Title : SW846 8260
QLast Update : Fri May 10 06:48:32 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.66 min Scan# 1832

Delta R.T. -0.00 min

Lab File: VN055567.D

Acq: 13 May 2019 10:04

Instrument :

MSVOA_N

ClientSampleId :


VN0513MBL01

Tot Ion:168 Resp: 789356

Ion Ratio Lower Upper




168 100

99 57.1 47.0 70.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO



#33

1,2-Dichloroethane-d4

Concen: 47.976 ug/l

RT: 8.03 min Scan# 1945

Delta R.T. -0.00 min

Lab File: VN055567.D


Acq: 13 May 2019 10:04

Tgt Ion: 65 Resp: 570202

Ion Ratio Lower Upper



65 100


67 54.0 0.0 109.6



Abundance Ion 64.90 (64.60 to 65.60): VNO

Ion 66.90 (66.60 to 67.60): VNO





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.58 min Scan# 2118

Delta R.T. -0.00 min

Lab File: VN055567.D

Acq: 13 May 2019 10:04

Instrument :

MSVOA_N

ClientSampleId :

VN0513MBL01


Tot Ion:114 Resp: 1321316

Ion Ratio Lower Upper

114 100

63 22.5 0.0 47.4




88 16.0 0.0 34.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN

Ion 87.90 (87.60 to 88.60): VN



#35

Dibromofluoromethane

Concen: 52.943 ug/l

RT: 7.59 min Scan# 1809

Delta R.T. -0.00 min

Lab File: VN055567.D

Acq: 13 May 2019 10:04


Tgt Ion:113 Resp: 451263

Ion Ratio Lower Upper

113 100

111 103.6 82.9 124.3



192 19.5 15.6 23.4




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 51.475 ug/l

RT: 10.09 min Scan# 2586

Delta R.T. -0.00 min

Lab File: VN055567.D

Acq: 13 May 2019 10:04

Instrument :

MSVOA_N

ClientSampleId :





VN0513MBL01

Tot Ion: 98 Resp: 1709747

Ion Ratio Lower Upper

98 100

100 63.3 50.4 75.6



#62

4-Bromofluorobenzene

Concen: 41.848 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. -0.00 min

Lab File: VN055567.D

Acq: 13 May 2019 10:04




Tgt Ion: 95 Resp: 468688


Ion Ratio Lower Upper

95 100

174 79.4 0.0 152.2

176 75.6 0.0 147.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. -0.00 min
Lab File: VN055567.D
Acq: 13 May 2019 10:04

Instrument :
MSVOA_N
ClientSampleId :
VN0513MBL01

Tot Ion:117 Resp: 1082583
Ion Ratio Lower Upper
117 100
82 58.5 48.0 72.0
119 32.0 26.0 39.0




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.35 min Scan# 3599
Delta R.T. 0.00 min
Lab File: VN055567.D
Acq: 13 May 2019 10:04

Tgt Ion:152 Resp: 272641
Ion Ratio Lower Upper
152 100
115 58.6 30.1 90.5
150 157.5 0.0 355.0



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

