

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051322\
 Data File : VN072472.D
 Acq On : 13 May 2022 12:55
 Operator : JC\MD
 Sample : VN0513WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0513WBL01

Quant Time: May 13 14:39:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 10 15:43:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	332452	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	499714	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	441226	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	168888	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	236945	46.532	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.060%
35) Dibromofluoromethane	8.022	113	164572	49.502	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.000%
50) Toluene-d8	10.439	98	619462	48.924	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.840%
62) 4-Bromofluorobenzene	12.727	95	199276	47.390	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051322\
Data File : VN072472.D
Acq On : 13 May 2022 12:55
Operator : JC\MD
Sample : VN0513WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0513WBL01

Quant Time: May 13 14:39:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
Quant Title : SW846 8260
QLast Update : Tue May 10 15:43:28 2022
Response via : Initial Calibration

