

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051322\
 Data File : VN072478.D
 Acq On : 13 May 2022 15:37
 Operator : JC\MD
 Sample : N2680-10
 Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B5-5-7

Quant Time: May 14 02:35:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 10 15:43:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	285454	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	433631	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	380735	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	158938	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	199154	45.550	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.100%
35) Dibromofluoromethane	8.016	113	136898	47.453	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.900%
50) Toluene-d8	10.439	98	516165	46.978	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	93.960%
62) 4-Bromofluorobenzene	12.727	95	176611	48.401	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.800%
Target Compounds						
25) 2-Butanone	7.351	43	19265	7.050	ug/l	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051322\
Data File : VN072478.D
Acq On : 13 May 2022 15:37
Operator : JC\MD
Sample : N2680-10
Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
B5-5-7

Quant Time: May 14 02:35:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
Quant Title : SW846 8260
QLast Update : Tue May 10 15:43:28 2022
Response via : Initial Calibration

