

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051322\
 Data File : VN072484.D
 Acq On : 13 May 2022 18:02
 Operator : JC\MD
 Sample : N2800-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-16

Quant Time: May 14 02:37:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 10 15:43:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	300857	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	450292	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	398378	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	146165	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	227304	49.327	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.660%
35) Dibromofluoromethane	8.022	113	154548	51.589	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.180%
50) Toluene-d8	10.439	98	573318	50.249	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.500%
62) 4-Bromofluorobenzene	12.727	95	179739	47.435	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.880%
Target Compounds						
						Qvalue
30) Chloroform	7.816	83	22663	3.446	ug/l	93
84) 1,2,4-Trimethylbenzene	13.363	105	2418	0.251	ug/l	100
95) Naphthalene	15.504	128	4085	3.724	ug/l	# 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051322\
Data File : VN072484.D
Acq On : 13 May 2022 18:02
Operator : JC\MD
Sample : N2800-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-16

Quant Time: May 14 02:37:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
Quant Title : SW846 8260
QLast Update : Tue May 10 15:43:28 2022
Response via : Initial Calibration

