

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051322\
 Data File : VN072488.D
 Acq On : 13 May 2022 19:39
 Operator : JC\MD
 Sample : N2800-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-15

Quant Time: May 14 02:38:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 10 15:43:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	305629	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	460112	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	396381	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	144401	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	229573	49.041	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.080%
35) Dibromofluoromethane	8.022	113	154826	50.578	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.160%
50) Toluene-d8	10.439	98	574026	49.238	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.480%
62) 4-Bromofluorobenzene	12.727	95	179693	46.411	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.820%
Target Compounds						
					Qvalue	
20) Methylene Chloride	5.104	84	2462	0.648	ug/l #	69
30) Chloroform	7.804	83	4868	0.729	ug/l #	67
68) m/p-Xylenes	11.951	106	1870	0.353	ug/l	85
69) o-Xylene	12.286	106	1409	0.271	ug/l	60
80) 1,3,5-Trimethylbenzene	13.057	105	3422	0.349	ug/l	92
84) 1,2,4-Trimethylbenzene	13.368	105	10243	1.075	ug/l	95
95) Naphthalene	15.498	128	11945	4.797	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051322\
Data File : VN072488.D
Acq On : 13 May 2022 19:39
Operator : JC\MD
Sample : N2800-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-15

Quant Time: May 14 02:38:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
Quant Title : SW846 8260
QLast Update : Tue May 10 15:43:28 2022
Response via : Initial Calibration

