

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051322\
 Data File : VN072489.D
 Acq On : 13 May 2022 20:03
 Operator : JC\MD
 Sample : N2800-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB050922

Quant Time: May 14 02:38:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 10 15:43:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	299908	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	444392	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	378128	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	138731	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	221540	48.228	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.460%
35) Dibromofluoromethane	8.016	113	149835	50.679	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.360%
50) Toluene-d8	10.439	98	545540	48.449	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.900%
62) 4-Bromofluorobenzene	12.727	95	171206	45.783	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.560%
Target Compounds						
16) Acetone	4.293	43	4503	2.475	ug/l	Qvalue # 70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051322\
Data File : VN072489.D
Acq On : 13 May 2022 20:03
Operator : JC\MD
Sample : N2800-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB050922

Quant Time: May 14 02:38:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
Quant Title : SW846 8260
QLast Update : Tue May 10 15:43:28 2022
Response via : Initial Calibration

