

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051322\
 Data File : VN072490.D
 Acq On : 13 May 2022 20:27
 Operator : JC\MD
 Sample : N2800-15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB051022

Quant Time: May 14 02:39:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 10 15:43:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	294508	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	442506	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	373428	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	137738	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	219889	48.746	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.500%
35) Dibromofluoromethane	8.022	113	150141	50.999	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.000%
50) Toluene-d8	10.439	98	544986	48.607	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.220%
62) 4-Bromofluorobenzene	12.727	95	170429	45.770	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.540%
Target Compounds						
16) Acetone	4.292	43	4618	2.585	ug/l	# 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051322\
Data File : VN072490.D
Acq On : 13 May 2022 20:27
Operator : JC\MD
Sample : N2800-15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB051022

Quant Time: May 14 02:39:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
Quant Title : SW846 8260
QLast Update : Tue May 10 15:43:28 2022
Response via : Initial Calibration

