

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051322\
 Data File : VN072492.D
 Acq On : 13 May 2022 21:15
 Operator : JC\MD
 Sample : N2666-01 10X
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WASTE

Quant Time: May 14 02:39:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 10 15:43:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	302342	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	449111	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	394864	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	150752	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	208108	44.939	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	89.880%
35) Dibromofluoromethane	8.016	113	141862	47.479	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.960%
50) Toluene-d8	10.439	98	542645	47.686	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.380%
62) 4-Bromofluorobenzene	12.727	95	180674	47.807	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.620%
Target Compounds						
					Qvalue	
39) Methylcyclohexane	9.451	83	2496	0.518	ug/l	91
69) o-Xylene	12.274	106	2852	0.550	ug/l	97
80) 1,3,5-Trimethylbenzene	13.057	105	24678	2.413	ug/l	97
84) 1,2,4-Trimethylbenzene	13.363	105	23325	2.344	ug/l	93
85) sec-Butylbenzene	13.492	105	5761	0.482	ug/l	98
86) p-Isopropyltoluene	13.615	119	8753	0.941	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051322\
Data File : VN072492.D
Acq On : 13 May 2022 21:15
Operator : JC\MD
Sample : N2666-01 10X
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WASTE

Quant Time: May 14 02:39:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
Quant Title : SW846 8260
QLast Update : Tue May 10 15:43:28 2022
Response via : Initial Calibration

