



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051322\
 Data File : VN072469.D
 Acq On : 13 May 2022 10:02
 Operator : JC\MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051022W.M
 Title : SW846 8260
 Last Update : Tue May 10 15:43:28 2022



AutoFind: Scans 1825, 1826, 1827; Background Corrected with Scan 1817

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.1	8756	PASS
75	95	30	60	55.5	23093	PASS
95	95	100	100	100.0	41592	PASS
96	95	5	9	6.7	2795	PASS
173	174	0.00	2	2.0	664	PASS
174	95	50	100	80.2	33339	PASS
175	174	5	9	7.6	2527	PASS
176	174	95	101	96.6	32200	PASS
177	176	5	9	6.7	2150	PASS