

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051325\
 Data File : VN086597.D
 Acq On : 13 May 2025 13:22
 Operator : JC\MD
 Sample : PB167932ZHE#16
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167932ZHE#16

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/14/2025
 Supervised By : Mahesh Dadoda 05/14/2025

Quant Time: May 14 01:43:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	171547	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	322142	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	321243	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	135518	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	122329	49.176	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	98.360%		
35) Dibromofluoromethane	8.165	113	94947	63.505	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	127.020%#		
50) Toluene-d8	10.565	98	417993	52.311	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	104.620%		
62) 4-Bromofluorobenzene	12.847	95	150408	51.607	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	103.220%		
Target Compounds						
16) Acetone	4.430	43	23031	21.564	ug/l	99
18) Methyl Acetate	5.042	43	3239	1.085	ug/l #	51
20) Methylene Chloride	5.283	84	4810	2.092	ug/l #	73
43) Isopropyl Acetate	8.688	43	25744m	2.751	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051325\
Data File : VN086597.D
Acq On : 13 May 2025 13:22
Operator : JC\MD
Sample : PB167932ZHE#16
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167932ZHE#16

Manual Integrations
APPROVED

Reviewed By : John Carlone 05/14/2025
Supervised By : Mahesh Dadoda 05/14/2025

Quant Time: May 14 01:43:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

