

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051325\
 Data File : VN086605.D
 Acq On : 13 May 2025 16:36
 Operator : JC\MD
 Sample : PB167964ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167964ZHE#03

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/14/2025
 Supervised By : Mahesh Dadoda 05/14/2025

Quant Time: May 14 01:47:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	160860	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	310136	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	306509	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	128783	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	115742	49.619	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.240%		
35) Dibromofluoromethane	8.165	113	92859	64.513	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	129.020%#		
50) Toluene-d8	10.565	98	401210	52.154	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	104.300%		
62) 4-Bromofluorobenzene	12.847	95	143278	51.064	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	102.120%		
Target Compounds						
16) Acetone	4.430	43	31606	34.265	ug/l	95
18) Methyl Acetate	5.036	43	4031	1.440	ug/l	92
20) Methylene Chloride	5.283	84	2619	1.214	ug/l #	81
43) Isopropyl Acetate	8.682	43	36224m	4.891	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051325\
Data File : VN086605.D
Acq On : 13 May 2025 16:36
Operator : JC\MD
Sample : PB167964ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167964ZHE#03

Quant Time: May 14 01:47:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/14/2025
Supervised By :Mahesh Dadoda 05/14/2025

