

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051325\
 Data File : VN086613.D
 Acq On : 13 May 2025 19:50
 Operator : JC\MD
 Sample : PB167964ZHE#11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167964ZHE#11

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/14/2025
 Supervised By : Mahesh Dadoda 05/14/2025

Quant Time: May 14 01:50:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	152681	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	291982	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	293471	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	119797	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	111005	50.138	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	= 100.280%		
35) Dibromofluoromethane	8.165	113	88766	65.504	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 131.000%#		
50) Toluene-d8	10.565	98	385566	53.237	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	= 106.480%		
62) 4-Bromofluorobenzene	12.847	95	136408	51.638	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery	= 103.280%		
Target Compounds						Qvalue
16) Acetone	4.430	43	31500	36.272	ug/l	98
18) Methyl Acetate	5.041	43	3735	1.405	ug/l #	81
20) Methylene Chloride	5.283	84	3329m	1.626	ug/l	
43) Isopropyl Acetate	8.688	43	33827m	4.836	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051325\
Data File : VN086613.D
Acq On : 13 May 2025 19:50
Operator : JC\MD
Sample : PB167964ZHE#11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167964ZHE#11

Quant Time: May 14 01:50:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 05/14/2025
Supervised By : Mahesh Dadoda 05/14/2025

