


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN051419\
Data File : VN055584.D
Acq On : 14 May 2019 9:02
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N051419W.M
Title : SW846 8260
Last Update : Wed May 15 07:28:47 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.9	21418	PASS
75	95	30	60	47.8	57224	PASS
95	95	100	100	100.0	119594	PASS
96	95	5	9	6.8	8084	PASS
173	174	0.00	2	1.0	965	PASS
174	95	50	100	77.3	92421	PASS
175	174	5	9	7.0	6495	PASS
176	174	95	101	99.4	91848	PASS
177	176	5	9	6.7	6126	PASS