

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051420\
 Data File : VN061405.D
 Acq On : 14 May 2020 20:25
 Operator : JC/MD
 Sample : L2591-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 3-11-27TH-AVE

Quant Time: May 15 07:40:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N051420W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 14 16:04:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.16	128	16045	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.55	114	95760	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.38	117	83014	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	7.99	65	45270	31.51	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	105.03%
60) 4-Bromofluorobenzene	12.38	95	39039	27.55	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	91.83%
63) Toluene-d8	10.06	98	120343	31.31	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	104.37%

Target Compounds					Ovalue
91) Naphthalene	15.11	128	3275	0.938 ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN051420\
Data File : VN061405.D
Acq On : 14 May 2020 20:25
Operator : JC/MD
Sample : L2591-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
3-11-27TH-AVE

Quant Time: May 15 07:40:12 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N051420W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu May 14 16:04:10 2020
Response via : Initial Calibration

