



Acrolein



$\text{Response} = 2.688\text{e-}002 * \text{Amt} + 4.204\text{e-}003$
 Coef of Det (r^2) = 0.992176 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N051421W.M
 Calibration Table Last Updated: Mon May 17 16:06:46 2021

Naphthalene


Response Ratio



$\text{Response} = 2.273\text{e}+000 * \text{Amt} - 1.479\text{e}-001$
 Coef of Det (r^2) = 0.996199 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N051421W.M
 Calibration Table Last Updated: Mon May 17 16:06:46 2021

Methylene Chloride

Response Ratio




$$\text{Response} = 5.502\text{e-}001 * \text{Amt} - 8.367\text{e-}004$$

Coef of Det (r^2) = 0.999417 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N051421W.M

Calibration Table Last Updated: Mon May 17 16:06:46 2021


Methyl methacrylate



Response = $3.343\text{e-}001 \times \text{Amt} - 1.299\text{e-}002$
Coef of Det (r^2) = 0.998361 Curve Fit: Linear
Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N051421W.M
Calibration Table Last Updated: Mon May 17 16:06:46 2021

4-Methyl-2-Pentanone

Response Ratio



$$\text{Response} = 4.108\text{e-}001 * \text{Amt} - 4.437\text{e-}002$$


Coef of Det (r^2) = 0.999517 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N051421W.M

Calibration Table Last Updated: Mon May 17 16:06:46 2021

Ethyl methacrylate

Response Ratio



$$\text{Response} = 5.690\text{e-}001 * \text{Amt} - 2.210\text{e-}002$$


Coef of Det (r^2) = 0.998249 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N051421W.M

Calibration Table Last Updated: Mon May 17 16:06:46 2021

2-Chloroethyl Vinyl ether

Response Ratio



$$\text{Response} = 9.356\text{e-}002 * \text{Amt} - 4.038\text{e-}002$$


Coef of Det (r^2) = 0.989829 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N051421W.M

Calibration Table Last Updated: Mon May 17 16:06:46 2021

2-Hexanone

Response Ratio




$$\text{Response} = 2.843\text{e-}001 * \text{Amt} - 1.246\text{e-}001$$

Coef of Det (r^2) = 0.997972 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N051421W.M


Calibration Table Last Updated: Mon May 17 16:06:46 2021

Bromoform



$\text{Response} = 3.970\text{e-}001 * \text{Amt} - 1.400\text{e-}002$
 Coef of Det (r^2) = 0.998641 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N051421W.M
 Calibration Table Last Updated: Mon May 17 16:06:46 2021


trans-1,4-Dichloro-2-butene



Response = $3.750 \times 10^{-1} \times \text{Amt} - 3.167 \times 10^{-2}$
Coef of Det (r^2) = 0.997953 Curve Fit: Linear
Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N051421W.M
Calibration Table Last Updated: Mon May 17 16:06:46 2021

4-Chlorotoluene

Response Ratio



$$\text{Response} = 2.739\text{e}+000 * \text{Amt} - 6.469\text{e}-002$$

Coef of Det (r^2) = 0.999119 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N051421W.M

Calibration Table Last Updated: Mon May 17 16:06:46 2021