

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051424\
 Data File : VN082117.D
 Acq On : 14 May 2024 15:48
 Operator : JC\MD
 Sample : VN0514WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0514WBL01

Quant Time: May 15 02:48:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue May 14 11:29:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.818	128	63918	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	338478	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	298123	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.577	65	169516	31.676	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	105.600%
60) 4-Bromofluorobenzene	12.847	95	121921	24.574	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	81.900%
63) Toluene-d8	10.565	98	417843	29.513	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	98.367%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051424\
Data File : VN082117.D
Acq On : 14 May 2024 15:48
Operator : JC\MD
Sample : VN0514WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0514WBL01

Quant Time: May 15 02:48:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue May 14 11:29:16 2024
Response via : Initial Calibration

