



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051424\
Data File : VN082113.D
Acq On : 14 May 2024 12:55
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Tue May 14 11:29:16 2024



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1844

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	26.7	19200	PASS
75	95	30	60	61.6	44315	FAIL*
95	95	100	100	100.0	71955	PASS
96	95	5	9	6.2	4485	PASS
173	174	0.00	2	1.0	569	PASS
174	95	50	100	76.8	55261	PASS
175	174	5	9	5.9	3279	PASS
176	174	95	101	95.9	53021	PASS
177	176	5	9	6.4	3417	PASS