

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051425\
 Data File : VN086617.D
 Acq On : 14 May 2025 11:10
 Operator : JC\MD
 Sample : VN0514MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0514MBL01

Quant Time: May 15 02:59:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	174321	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	332328	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	323196	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	134251	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	110538	43.729	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.460%
35) Dibromofluoromethane	8.165	113	91271	59.176	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	118.360%
50) Toluene-d8	10.565	98	387539	47.013	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.020%
62) 4-Bromofluorobenzene	12.847	95	136354	45.351	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	90.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051425\
Data File : VN086617.D
Acq On : 14 May 2025 11:10
Operator : JC\MD
Sample : VN0514MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0514MBL01

Quant Time: May 15 02:59:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

