

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051425\
 Data File : VN086626.D
 Acq On : 14 May 2025 14:56
 Operator : JC\MD
 Sample : Q2014-01
 Misc : 1.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MOO-25-0134

Quant Time: May 15 03:04:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	182310	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.094	114	350795	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	324073	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	136078	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	113447	42.913	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.820%
35) Dibromofluoromethane	8.159	113	92699	56.937	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	113.880%
50) Toluene-d8	10.565	98	397950	45.735	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.460%
62) 4-Bromofluorobenzene	12.847	95	133648	42.111	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	84.220%
Target Compounds						
					Qvalue	
18) Methyl Acetate	5.036	43	10417	3.283	ug/l	91
86) p-Isopropyltoluene	13.723	119	8514	0.943	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN051425\
Data File : VN086626.D
Acq On : 14 May 2025 14:56
Operator : JC\MD
Sample : Q2014-01
Misc : 1.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MOO-25-0134

Quant Time: May 15 03:04:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

